



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No 9410
 Dichloromethane
 'BAKER HPLC ANALYZED'
 for use in High Performance Liquid Chromatography

Lot No 1326601861
Release Date 27-Sep-2013

Test	Specification	Actual value
Assay (by GC)	min. 99.8 %	100.0 %
Chloride (Cl)	max. 10 ppm	< 10 ppm
Density (g/ml) at 25°C	1.315-1.321	1.320
Preservative (Amylene) (ppm)	about 20	passes test
Residue after Evaporation (in ppm)	max. 2	< 2
Titration Acid (meq/g)	max. 0.0003	0.00004
Water (H ₂ O)	max. 0.02 %	0.004 %
Ultraviolet Absorbance (1.00-cm path vs water):		
at 254 nm	max. 0.01	< 0.01
at 280 nm	max. 0.01	< 0.01
at 350 nm	max. 0.01	< 0.01
UV Cut-off, nm	max. 233	229

Use Before Date: 2015-09

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials B.V.

Avantor Performance Materials B.V. Teugseweg 20 7418 AM Deventer The Netherlands Tel:+31 570887500

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

Dr. J. Mittendorf
 QA & RA Manager