



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

## Certificate of Analysis

**Product No** 7307-03  
 Amino (NH<sub>2</sub>)  
 BAKERBOND spe™  
**Lot No** 1414603016  
**Release Date** 26-May-14

Test	Specification	Actual value
<i>The following results are based on the bulk bonded phase:</i>		
Carbon (C)(%)	For Information Only	7
Nitrogen (N)(%)	For Information Only	2.71
Surface Coverage (I/g/m <sup>2</sup> ), Based on %C	For Information Only	134
Surface Coverage (I/g/m <sup>2</sup> ), Based on %N	For Information Only	45
<i>The following results are based on the silica gel raw material:</i>		
Particle Size Distribution: <31 microns, wt%	max. 17 %	12 %
Particle Size Distribution: APD, microns	47 - 61	53
Specific Surface Area, m <sup>2</sup> /g	For Information Only	512
<i>The following results are based on the solid phase extraction columns:</i>		
Extraction Residue	max. 0.2 %	passes test

For Laboratory, Research or Manufacturing Use

Use Before Date: 2019-06

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at [avantor.emea@avantormaterials.com](mailto:avantor.emea@avantormaterials.com)

The information in this document is property of Avantor Performance Materials B.V.

Avantor Performance Materials B.V. Teugseweg 20 7418 AM Deventer The Netherlands Tel:+31 570687500

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

[JTBaker\\_REACH\\_USERS@avantormaterials.com](mailto:JTBaker_REACH_USERS@avantormaterials.com) indicating the JTBaker product number you purchase.

Dr. J Mittendorf  
 QA & RA Manager